

Message

From: Daly, Eric [Daly.Eric@epa.gov]
Sent: 3/29/2016 8:32:43 PM
To: Nguyen, Lyndsey [Nguyen.Lyndsey@epa.gov]; Gaughan, Daniel [Gaughan.Daniel@epa.gov]; Bernard Nwosu (Ben.Nwosu@WestonSolutions.com) [Ben.Nwosu@WestonSolutions.com]
Subject: CRU Equipment
Importance: High

Good Afternoon Dan:

Just some clarification. Ben just informed me that the Ludlum 2241 with the 10 foot cable was rented. We just need to make sure of a few things. Can you ensure that the instrument is calibrated with the long cable? If not, what specific cable was it calibrated with and do you have it with the unit? Lyndsey will need the cable used for calibration to do a source check IF it was not calibrated with long cable. Lyndsey, I assume you do not need to bring your unit and cable but you all can figure out if that is needed.

Please note, we do not need the rented long cable Ludlum unit for HTC site work. Thanks

- Two (2) RSS-131 HPIC (**Will be transported to Edison on 03/28/16**)
- Two (2) Ludlum 2241, with 3"x3" NaI probes (**Please note we are using a more sensitive 3x3 probe than used in the past, will be transported to Edison on 03/28/16**)
- One (1) Ludlum 2241 with 44-10 2"x2" NaI Probe, with 5-6 foot cable and 10 foot cable (**Lyndsey will provide instrumentation and cables**)

Regards,

Eric

"We must, indeed, all hang together, or assuredly we shall all hang separately", Benjamin Franklin

Eric M. Daly
On-Scene Coordinator/Radiological Response Specialist
US Environmental Protection Agency- Region II
ERRD/RPB/PPS
2890 Woodbridge Avenue
Edison, NJ 08837
daly.eric@epa.gov
732-321-4350